

**3701-83-55 Quality assessment and performance improvement-
freestanding diagnostic imaging centers.**

- (A) As part of the quality assessment and performance improvement program required under paragraph (C) of rule 3701-83-12 of the Administrative Code, each freestanding or mobile diagnostic imaging center shall:
 - (1) Establish and maintain a clinical image quality control program to monitor and document images repeated due to poor image quality;
 - (2) Monitor and evaluate the accuracy of image interpretations by:
 - (a) Establishing a clinical image review program;
 - (b) Establishing policies and procedures and the semiannual audit of a random, representative sample of total clinical images performed at the center;
 - (c) Provide for and implement the semiannual audit conducted by a radiologist or by a physician or chiropractor as identified under paragraphs (C)(2), (D)(2), (E)(2), and (F)(2) of rule 3701-83-52 of the Administrative Code:
 - (i) Diagnostic imaging centers under the operation of more than one physician or chiropractic radiologist may conduct their own internal semiannual audit of a random sample of clinical images.
 - (ii) Diagnostic imaging centers operated by a single physician or chiropractic radiologist shall provide for the external semiannual audit of a random sample of clinical images. External audits shall not be conducted by an individual who is associated with, or has a financial interest in the center.
 - (3) Monitor and evaluate any problems associated with sedation, the administration of contrast agents, and problems identified as part of the semiannual audit of a random sample of clinical images.
- (B) As part of the quality assessment and performance improvement program required under paragraph (A) of rule 3701-83-12 of the Administrative Code, each freestanding or mobile diagnostic imaging center shall report to the director:
 - (1) The number, type, and age, of diagnostic imaging equipment, including magnetic strength when applicable, and whether diagnostic imaging equipment is fixed or mobile;
 - (2) If diagnostic imaging equipment is mobile, a list of the locations where the service is delivered, including contracted sites for which the service is currently not operational;
 - (3) The number of procedures performed by CPT code;
 - (4) The number of repeat procedures taken which were conducted in order to obtain a sufficient image relating to the patient's image order;
 - (5) The number and type of complications associated with sedation and the administration of contrast agents;

- (6) The number of patients who required hospitalization, as a result of a complication, within twenty-four hours of a procedure;
- (7) The number of diagnostic procedures performed and the number that revealed negative findings; and
- (8) The number of clinical images audited by the center or an external individual, and the number of cases in which there was variance between audited findings and original findings.

Effective: 07/01/2016

Five Year Review (FYR) Dates: 02/16/2016 and 02/15/2021

CERTIFIED ELECTRONICALLY

Certification

04/22/2016

Date

Promulgated Under: 119.03
Statutory Authority: 3702.13; 3702.30
Rule Amplifies: 3702.12, 3702.13; 3702.30
Prior Effective Dates: 1/13/1996, 6/27/96, 6/1/06, 4/24/11