

3701:1-58-23 **Possession, use, and calibration of instruments used to measure the activity of unsealed radioactive material.**

- (A) For direct measurements performed in accordance with rule 3701:1-58-25 of the Administrative Code, a licensee shall possess and use instrumentation to measure the activity of unsealed radioactive material before it is administered to each patient or human research subject.
- (B) A licensee shall calibrate the instrumentation required in paragraph (A) of this rule in accordance with nationally recognized standards or the manufacturer's instructions.
- (C) A licensee shall retain a record of each instrument calibration required by this rule in accordance with rule 3701:1-58-77 of the Administrative Code.

Five Year Review (FYR) Dates: 06/12/2015 and 06/01/2020

CERTIFIED ELECTRONICALLY

Certification

06/12/2015

Date

Promulgated Under: 119.03
Statutory Authority: 3748.02, 3748.04
Rule Amplifies: 3748.04
Prior Effective Dates: 8/15/2005