

3701:1-58-85 **Records of molybdenum-99, strontium-82, and strontium-85 concentrations.**

A licensee shall maintain a record of the molybdenum-99 concentration or strontium-82 and strontium-85 concentration tests required by paragraph (B) of rule 3701:1-58-35 of the Administrative Code for three years. The record must include:

- (A) For each measured elution of technetium-99m, the ratio of the measures expressed as kilobecquerel (microcuries) of molybdenum-99 per megabecquerel (millicurie) of technetium-99m, the time and date of the measurement, and the name of the individual who made the measurement; or
- (B) For each measured elution of rubidium-82, the ratio of the measures expressed as kilobecquerel (microcuries) of strontium-82 per megabecquerel (millicurie) of rubidium-82, kilobecquerel (microcuries) of strontium-85 per megabecquerel (millicurie) of rubidium-82, the time and date of the measurement, and the name of the individual who made the measurement.

Five Year Review (FYR) Dates: 06/12/2015 and 06/01/2020

CERTIFIED ELECTRONICALLY

\_\_\_\_\_  
Certification

06/12/2015

\_\_\_\_\_  
Date

Promulgated Under: 119.03  
Statutory Authority: 3748.02, 3748.04  
Rule Amplifies: 3748.04  
Prior Effective Dates: 8/15/2005, 10/4/10