OHIO DEPARTMENT OF HEALTH

ALCOHOL AND DRUG TESTING

APPROVAL OF INSTRUMENT CHECK SOLUTION, OAC 3701-53-04

Lot or Batch # ODH-0020 Target Value 0.099 g/210L

This instrument check solution is to be used in a simulator for the instrument check of approved evidential breath testing instruments, as defined in OAC 3701-53-02.

This instrument check solution contains **1.202 mg/mL** ethyl alcohol in purified water. When used according to instrument check forms, it will produce a result at or within ±0.005 g/210L of the target value in the approved breath testing instrument, when that instrument is in proper working condition.

This instrument check solution shall not be used more than three months after its date of first use or more than one year after its date of manufacture, May 14, 2013. Whichever comes first.

This solution expires May 14, 2014

Approved this 12th day of June, 2013

Director of Health

LITHO IN U.S.A.