

**Radiation-generating Equipment Committee (REC)**

**June 8, 2012**

**Minutes**

**MEMBERS PRESENT**

Lawrence Osher, Chair  
John Dukes  
Chuck Wissuchek  
Bryon Murray

**MEMBERS ABSENT**

Kathryn Gardner  
Thomas Hangartner  
Brenda Johnson  
David Kimberly  
Bradley Kraai  
Nina Mayr  
Terry Yates  
Nina Kowalczyk  
Jill Paessun  
Ruth Hackworth  
Kerry Krugh, Vice Chair

**GUESTS**

Paul Geis, University Hospitals Cleveland  
Jim Burn, Flower & Sylvania Hospital  
Susan Suchan, Mt. Carmel  
Ron Droege, Atrium Medical Center  
Andy Anderson, MRP, Inc.  
Jeremy Donaghue, Akron General Medical Center

**ODH ATTENDEES**

James Castle  
Dwight Leeseberg

\*\*\*\*\*

The Radiation-Generating Equipment Committee (REC) meeting was called to order by chairperson, Larry Osher at 10:10 a.m. The meeting was held at the Ohio Department of Health (ODH) in the Basement Training Room A at 35 Chestnut Street, Columbus, Ohio. The sign-in sheet serves as the official record of attendance.

**Past Minutes:**

The committee reviewed the May 4, 2012 minutes. The minutes were approved with a correction to Bryon Murray's name and the addition of two guests, Paul Johnson and Dave Jordon who did not sign the sign-in sheet. Jack Dukes made the motion to accept the minutes with the changes. Bryon Murray seconded the motion; the members present unanimously approved the minutes with the changes.

**Current Business:**

Jim Castle provided the rule status log and explained that the only rule due for 5 year review is 3701:1-38-04. REC suggested that the status log be sorted by review date instead of rule number.

The REC and ODH covered all public comments to the 12/11/2011 draft therapy rules 67-05, 67-06, 67-08 and 67-09 and developed new drafts rules from these comments. The REC suggested moving some language from the appendix of rule 3701:1-67-08 into the rule itself. A topic of concern regarding neutron shielding calculations to include neutron spectra as a function of neutron fluence rate in rule 3701:1-67-08 is under debate.

At the July 13, 2012 meeting, REC plans to review and update definitions of rule 3701:1-67-01 to match the newly drafted therapy rules. Review language moved from the appendix of rule 3701:1-67-08 into the rule itself. Finish the discussion regarding the assumptions for neutron shielding calculation to include neutron spectra as a function of neutron fluence rate. Approve Chapter 3701:1-67 revisions to go out for public comment.

**Future Meeting Dates:**

Friday, July 13, 2012

**Adjourn:**

The meeting was adjourned at 4:00pm.