

**3701-72-04 Licensee authorizations and requirements.**

- (A) Each radiologic licensee shall operate ionizing radiation-generating equipment or use radiopharmaceuticals in accordance with the safe operating procedures and instruction of workers required under the rules adopted pursuant to Chapter 3748. of the Revised Code. The director may at any time during business hours evaluate operator competency in regard to the safe operation of ionizing radiation-generating equipment or use of radioactive sources to determine compliance with Chapter 4773. of the Revised Code or any rules adopted pursuant to Chapter 3748. of the Revised Code.
- (B) Radiographers, nuclear medicine technologists, radiation therapists, and general x-ray machine operators who are registered nurses providing occupational health nursing services in an industrial workplace shall work under general supervision.
- (C) General x-ray machine operators, other than those registered nurses who provide occupational health nursing services in an industrial workplace, shall work under the direct supervision of a licensed practitioner.
- (D) Provided that all radiation safety training required by rules adopted pursuant to Chapter 3748. of the Revised Code have been completed, registered nurses, registered cardiovascular invasive specialists, and general x-ray machine operators licensed prior to September 30, 1995, who have been continuously licensed in accordance with this chapter to practice in the area of fluoroscopy pursuant to the conditional license requirements of section 4773.05 of the Revised Code, may perform, under the direct supervision of a licensed practitioner present in the room, the following tasks of a cardiac catheterization fluoroscopic procedure:
  - (1) Move or position the patient during exposure;
  - (2) Move or position the fluoroscopic unit;
  - (3) Reset the timer after the audible signal; and
  - (4) Develop the radiographic image.
- (E) The following fluoroscopy tasks shall be performed only by a licensed practitioner or an Ohio licensed radiographer:
  - (1) Adjust or set technique factors for the x-ray procedure;
  - (2) Activate the switch to expose the patient; or
  - (3) Assure adequate radiation protection to the patient and individuals in the procedure room from unnecessary radiation.
- (F) General x-ray machine operators licensed after August 1, 2008, shall complete one or as many of the following competency-based clinical training modules from an accredited competency-based clinical training program needed to meet the requirements to work within the scope of practice of their supervising practitioner:
  - (1) Chest and abdomen,
  - (2) Extremities,
  - (3) Skull and sinus,

- (4) Spine,
  - (5) Podiatric radiography, or
  - (6) Bone densitometry.
- (G) Effective August 1, 2008, general x-ray machine operators changing employment to a different licensed practitioner within a new area of clinical practice, shall, prior to practicing in the new area, complete the appropriate competency-based clinical training module as required by paragraph (F) of this rule and submit evidence of completion of the clinical training to the department.
- (H) General x-ray machine operators initially licensed prior to August 1, 2008, who want to perform standard, diagnostic radiologic procedures using digital equipment, shall complete the corresponding competency-based clinical training module as listed in paragraph (F) of this rule and submit evidence of completion of the clinical training to the department.
- (I) A licensee shall post his or her license or a facsimile thereof at a conspicuous location in his or her work area at each place of employment, or make his or her license available immediately upon the request of any person.
- (J) Each licensee shall notify the department in writing within thirty days if:
- (1) There is a change in the licensee's name or address; or
  - (2) In the case of the general x-ray machine operator, there is a change of employment or a change in the licensed practitioner providing direct supervision.
- (K) Positron emission tomography-computed tomography and single photon emission computed tomography-computed tomography fusion procedures shall be performed by a licensed nuclear medicine technologist or a licensed radiographer who has obtained training and demonstrated competency with the equipment and procedures. Radiopharmaceuticals shall be injected only by a licensed nuclear medicine technologist or other competent licensed individuals acting within their scope of practice as authorized by the Revised Code. If the computed tomography scanner is operated for any other diagnostic radiographic procedures, the computed tomography scanner shall be operated by a licensed radiographer.
- (L) Radiographers or radiation therapists, who have obtained training and demonstrated competency with the computerized tomography (CT) equipment and CT protocols, may operate CT units for radiation therapy simulation procedures. CT simulation procedures shall be overseen by a radiation therapist. If the CT equipment is operated for any other diagnostic radiographic procedures, the CT equipment shall be operated by a radiographer.

Effective: 04/01/2012

R.C. 119.032 review dates: 12/30/2011 and 04/01/2017

CERTIFIED ELECTRONICALLY

---

Certification

03/22/2012

---

Date

Promulgated Under: 119.03  
Statutory Authority: 4773.08  
Rule Amplifies: 4773.03, 4773.06, 4773.08, 4773.99  
Prior Effective Dates: 10/16/1995, 10/13/97, 9/4/00, 5/17/04, 8/1/07