

3701:1-52-15 **Power failures.**

- (A) If electrical power at a panoramic irradiator is lost for longer than ten seconds, the sources must automatically return to the shielded position.
- (B) The lock on the door of the radiation room of a panoramic irradiator may not be deactivated by a power failure.
- (C) During a power failure, the area of any irradiator where sources are located may be entered only when using an operable and calibrated radiation survey meter.

Five Year Review (FYR) Dates: 10/25/2016 and 10/15/2021

CERTIFIED ELECTRONICALLY

Certification

10/25/2016

Date

Promulgated Under: 119.03
Statutory Authority: 3748.04
Rule Amplifies: 3748.04
Prior Effective Dates: 10/20/2002