

3701:1-58-100 **Records of five-year inspection for teletherapy and gamma stereotactic radiosurgery units.**

- (A) A licensee shall maintain a record of the five-year inspections for teletherapy and gamma stereotactic radiosurgery units required by rule 3701:1-58-69 of the Administrative Code for the duration of use of the unit.
- (B) The record must contain:
- (1) The inspector's radioactive materials license number;
 - (2) The date of inspection;
 - (3) The manufacturer's name and model number and serial number of both the treatment unit and source;
 - (4) A list of components inspected and serviced, and the type of service; and
 - (5) The signature of the inspector.

Five Year Review (FYR) Dates: 06/12/2015 and 06/01/2020

CERTIFIED ELECTRONICALLY

Certification

06/12/2015

Date

Promulgated Under: 119.03
Statutory Authority: 3743.02, 3748.04
Rule Amplifies: 3748.04
Prior Effective Dates: 8/15/2005