

3701:1-58-24

Calibration of survey instruments.

- (A) A licensee shall calibrate the survey instruments used to show compliance with this chapter and Chapter 3701:1-38 of the Administrative Code before first use, annually, and following a repair that affects the calibration. A licensee shall:
- (1) Calibrate all scales with readings up to ten millisievert (one thousand millirem) per hour with a radiation source;
 - (2) Calibrate two separate readings on each scale or decade that will be used to show compliance; and
 - (3) Conspicuously note on the instrument the date of calibration.
- (B) A licensee may not use survey instruments if the difference between the indicated exposure rate and the calculated exposure rate is more than twenty - per cent.
- (C) A licensee shall retain a record of each survey instrument calibration in accordance with rule 3701:1-58-78 of the Administrative Code.

Five Year Review (FYR) Dates: 06/12/2015 and 06/01/2020

CERTIFIED ELECTRONICALLY

Certification

06/12/2015

Date

Promulgated Under: 119.03
Statutory Authority: 3748.02, 3748.04
Rule Amplifies: 3748.04
Prior Effective Dates: 8/15/2005, 10/4/10