

3701:1-58-26 **Authorization for calibration, transmission, and reference sources.**

Any person authorized by rule 3701:1-58-06 of the Administrative Code for medical use of radioactive material may receive, possess, and use any of the following radioactive material for check, calibration, transmission, and reference use:

- (A) Sealed sources, not exceeding 1.11 gigabecquerels (thirty millicuries) each, manufactured and distributed by a person licensed under rule 3701:1-46-44 of the Administrative Code or equivalent United States nuclear regulatory commission, or agreement state regulations.
- (B) Sealed sources, not exceeding 1.11 gigabecquerels (thirty millicuries) each, redistributed by a licensee authorized to redistribute the sealed sources manufactured and distributed by a person licensed under rule 3701:1-46-44 of the Administrative Code or equivalent United States nuclear regulatory commission or agreement state regulations, providing the redistributed sealed sources are in the original packaging and shielding and are accompanied by the manufacturer's approved instructions.
- (C) Any radioactive material with a half-life not longer than one hundred twenty days in individual amounts not to exceed 0.56 gigabecquerels (fifteen millicuries).
- (D) Any radioactive material with a half-life longer than one hundred twenty days in individual amounts not to exceed the smaller of 7.4 megabecquerels (two hundred microcuries) or one thousand times the quantities in appendix A to rule 3701:1-40-17 of the Administrative Code.
- (E) Technetium-99m in amounts as needed.

R.C. 119.032 review dates: 07/31/2013 and 12/17/2017

CERTIFIED ELECTRONICALLY

\_\_\_\_\_  
Certification

07/31/2013

\_\_\_\_\_  
Date

Promulgated Under: 119.03  
Statutory Authority: 3748.02, 3748.04  
Rule Amplifies: 3748.04  
Prior Effective Dates: 8/15/2005, 12/22/08