

3701:1-58-49

**Decay of strontium-90 sources for ophthalmic treatments.**

- (A) Only an authorized medical physicist shall calculate the activity of each strontium-90 source that is used to determine the treatment times for ophthalmic treatments. The decay must be based on the activity determined under rule 3701:1-58-48 of the Administrative Code.
- (B) A licensee shall retain a record of the activity of each strontium-90 source in accordance with rule 3701:1-58-90 of the Administrative Code.

Five Year Review (FYR) Dates: 06/12/2015 and 06/01/2020

CERTIFIED ELECTRONICALLY

\_\_\_\_\_  
Certification

06/12/2015

\_\_\_\_\_  
Date

Promulgated Under: 119.03  
Statutory Authority: 3748.02, 3748.04  
Rule Amplifies: 3748.04  
Prior Effective Dates: 8/15/2005