

3701:1-58-50 **Therapy-related computer systems for manual brachytherapy.**

The licensee shall perform acceptance testing on the treatment planning system of therapy-related computer systems in accordance with published protocols accepted by nationally recognized bodies. At a minimum, the acceptance testing must include, as applicable, verification of:

- (A) The source-specific input parameters required by the dose calculation algorithm;
- (B) The accuracy of dose, dwell time, and treatment time calculations at representative points;
- (C) The accuracy of isodose plots and graphic displays; and
- (D) The accuracy of the software used to determine sealed source positions from radiographic images.

Five Year Review (FYR) Dates: 06/12/2015 and 06/01/2020

CERTIFIED ELECTRONICALLY

Certification

06/12/2015

Date

Promulgated Under: 119.03
Statutory Authority: 3748.02, 3748.04
Rule Amplifies: 3748.04
Prior Effective Dates: 8/15/2005