

3701:1-58-54

Training for use of sealed sources for diagnosis.

Except as provided in rule 3701:1-58-21 of the Administrative Code, the licensee shall require the authorized user of a diagnostic sealed source for use in a device authorized under rule 3701:1-58-53 of the Administrative Code to be a physician, dentist, or podiatrist who:

- (A) Is certified by a specialty board whose certification process includes all of the requirements in paragraphs (B) and (C) of this rule and whose certification has been recognized by the director, United States nuclear regulatory commission, or an agreement state. The names of board certifications which have been recognized by the director, United States nuclear regulatory commission, or an agreement state will be posted on the United States nuclear regulatory commission's web page at www.nrc.gov; or
- (B) Has completed eight hours of classroom and laboratory training in basic radionuclide handling techniques specifically applicable to the use of the device. The training must include:
 - (1) Radiation physics and instrumentation;
 - (2) Radiation protection;
 - (3) Mathematics pertaining to the use and measurement of radioactivity; and
 - (4) Radiation biology; and
- (C) Has completed training in the use of the device for the uses requested.

R.C. 119.032 review dates: 07/31/2013 and 12/17/2017

CERTIFIED ELECTRONICALLY

Certification

07/31/2013

Date

Promulgated Under: 119.03
Statutory Authority: 3748.02, 3748.04
Rule Amplifies: 3748.04
Prior Effective Dates: 8/15/2005, 12/22/08