

3701:1-58-78

Records of radiation survey instrument calibrations.

A licensee shall maintain a record of radiation survey instrument calibrations required by rule 3701:1-58-24 of the Administrative Code for three years. The record must include the model and serial number of the instrument, the date of the calibration, the results of the calibration, and the name of the individual who performed the calibration.

Five Year Review (FYR) Dates: 06/12/2015 and 06/01/2020

CERTIFIED ELECTRONICALLY

Certification

06/12/2015

Date

Promulgated Under: 119.03
Statutory Authority: 3743.02, 3748.04
Rule Amplifies: 3748.04
Prior Effective Dates: 8/15/2005