

3701:1-58-89 **Records of calibration measurements of brachytherapy sources.**

- (A) A licensee shall maintain a record of the calibrations of brachytherapy sources required by rule 3701:1-58-48 of the Administrative Code for three years after the last use of the source.
- (B) The record must include:
- (1) The date of the calibration;
  - (2) The manufacturer's name, model number, and serial number for the source and the instruments used to calibrate the source;
  - (3) The source output or activity;
  - (4) The source positioning accuracy within the applicators; and
  - (5) The name of the individual, the source manufacturer, or the calibration laboratory that performed the calibration.

Five Year Review (FYR) Dates: 06/12/2015 and 06/01/2020

CERTIFIED ELECTRONICALLY

---

Certification

06/12/2010

---

Date

Promulgated Under: 119.03  
Statutory Authority: 3743.02, 3748.04  
Rule Amplifies: 3748.04  
Prior Effective Dates: 8/15/2005