

3701:1-58-93 **Records of dosimetry equipment used with remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.**

- (A) A licensee shall retain a record of the calibration, intercomparison, and comparisons of its dosimetry equipment done in accordance with rule 3701:1-58-60 of the Administrative Code for the duration of the license.
- (B) For each calibration, intercomparison, or comparison, the record must include:
- (1) The date;
 - (2) The manufacturer's name, model numbers and serial numbers of the instruments that were calibrated, intercompared, or compared as required by paragraphs (A) and (B) of rule 3701:1-58-60 of the Administrative Code;
 - (3) The correction factor that was determined from the calibration or comparison or the apparent correction factor that was determined from an intercomparison; and
 - (4) The names of the individuals who performed the calibration, intercomparison, or comparison.

Five Year Review (FYR) Dates: 06/12/2015 and 06/01/2020

CERTIFIED ELECTRONICALLY

Certification

06/12/2015

Date

Promulgated Under: 119.03
Statutory Authority: 3743.02, 3748.04
Rule Amplifies: 3748.04
Prior Effective Dates: 8/15/2005