

3701:1-58-94 **Records of full calibrations for teletherapy units, remote afterloader units, and gamma stereotactic radiosurgery units.**

- (A) A licensee shall maintain a record of the teletherapy unit, remote afterloader unit, and gamma stereotactic radiosurgery unit full calibrations required by rules 3701:1-58-61 to 3701:1-58-63 of the Administrative Code for three years.
- (B) The record must include:
- (1) The date of the calibration;
 - (2) The manufacturer's name, model number, and serial number of the teletherapy, remote afterloader, and gamma stereotactic radiosurgery unit(s), the source(s), and the instruments used to calibrate the unit(s);
 - (3) The results and an assessment of the full calibrations;
 - (4) The results of the autoradiograph required for low dose-rate remote afterloader units; and
 - (5) The signature of the authorized medical physicist who performed the full calibration.

Five Year Review (FYR) Dates: 06/12/2015 and 06/01/2020

CERTIFIED ELECTRONICALLY

Certification

06/12/2015

Date

Promulgated Under: 119.03
Statutory Authority: 3748.02, 3748.04
Rule Amplifies: 3748.04
Prior Effective Dates: 8/15/2005, 10/4/10