

3701:1-58-95

Records of periodic spot-checks for teletherapy units.

- (A) A licensee shall retain a record of each periodic spot-check for teletherapy units required by rule 3701:1-58-64 of the Administrative Code for three years.
- (B) The record must include:
- (1) The date of the spot-check;
 - (2) The manufacturer's name, model number, and serial number of the teletherapy unit, source and instrument used to measure the output of the teletherapy unit;
 - (3) An assessment of timer linearity and constancy;
 - (4) The calculated on-off error;
 - (5) A determination of the coincidence of the radiation field and the field indicated by the light beam localizing device;
 - (6) The determined accuracy of each distance measuring and localization device;
 - (7) The difference between the anticipated output and the measured output;
 - (8) Notations indicating the operability of each entrance door electrical interlock, each electrical or mechanical stop, each source exposure indicator light, and the viewing and intercom system and doors; and
 - (9) The name of the individual who performed the periodic spot-check and the signature of the authorized medical physicist who reviewed the record of the spot-check.
- (C) A licensee shall retain a copy of the procedures required by paragraph (B) of rule 3701:1-58-64 of the Administrative Code until the licensee no longer possesses the teletherapy unit.

Five Year Review (FYR) Dates: 06/12/2015 and 06/01/2020

CERTIFIED ELECTRONICALLY

Certification

06/12/2015

Date

Promulgated Under: 119.03
Statutory Authority: 3743.02, 3748.04
Rule Amplifies: 3748.04

Prior Effective Dates: 8/15/2005